

Datasheet: MCA929PE

Description:	MOUSE IgG2a NEGATIVE CONTROL:RPE
Specificity:	MOUSE IgG2a NEGATIVE CONTROL
Format:	RPE
Product Type:	Negative/Isotype Control
Isotype:	IgG2a
Quantity:	100 TESTS

Product Details

Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-rad-antibodies.com/protocols.

	Yes	No	Not Determined	Suggested Dilution
Flow Cytometry	■			*

Where this antibody has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. * It is recommended that the user dilutes the antibody for use in their own system to a concentration equivalent to their test reagent.

Target Species	Negative Control		
Product Form	Purified IgG conjugated to R. Phycoerythrin (RPE) - lyophilized		
Reconstitution	Reconstitute with 1.0 ml distilled water		
Max Ex/Em	Fluorophore	Excitation Max (nm)	Emission Max (nm)
	RPE 488nm laser	496	578
Preparation	Purified IgG prepared by affinity chromatography on Protein G from tissue culture supernatant		
Buffer Solution	Phosphate buffered saline		
Preservative Stabilisers	0.09% Sodium Azide		
	1%	Bovine Serum Albumin	
	5%	Sucrose	
Immunogen	Activated rat T-helper cells.		
Fusion Partners	Spleen cells from immunised BALB/c mice were fused with cells of the NS1 mouse myeloma cell line.		
Specificity	Mouse IgG2a negative control antibody, clone OX-34 is suitable for use as a negative control reagent for the measurement of non-specific binding of mouse monoclonal antibodies of isotype IgG2a to human tissue.		

Clone MRC OX-34 recognises a rat cell surface marker, and therefore cannot be used as a negative control in this species.

This product is routinely tested in flow cytometry on rat splenocytes to confirm antibody activity and on human whole blood to test for suitability as a negative control.

Test results have shown that MCA929 is also suitable for use as a negative control with bovine, ovine, porcine, equine, canine, lapine and guinea-pig tissues.

This antibody may not be suitable for intracellular staining on some cell types.

Flow Cytometry	Use 10ul of the suggested working dilution to label 10 ⁶ cells in 100ul.
-----------------------	---

References	1. Avigdor, A. <i>et al.</i> (2004) CD44 and hyaluronic acid cooperate with SDF-1 in the trafficking of human CD34+ stem/progenitor cells to bone marrow. Blood. 103 (8): 2981-9. 2. Kamble, N.M. <i>et al.</i> (2016) Interaction of a live attenuated <i>Salmonella gallinarum</i> vaccine candidate with chicken bone marrow-derived dendritic cells. Avian Pathol. Jan 26:1-24. [Epub ahead of print] 3. Wattegedera, S.R. <i>et al.</i> (2017) Enhancing the toolbox to study IL-17A in cattle and sheep. Vet Res. 48 (1): 20.
-------------------	---

Storage	Prior to reconstitution store at +4°C. Following reconstitution store at +4°C.
----------------	--

DO NOT FREEZE.

This product should be stored undiluted. This product is photosensitive and should be protected from light. Should this product contain a precipitate we recommend microcentrifugation before use.

Shelf Life	12 months from date of reconstitution.
-------------------	--

Health And Safety Information	Material Safety Datasheet documentation #10075 available at: 10075: https://www.bio-rad-antibodies.com/uploads/MSDS/10075.pdf
--------------------------------------	--

Regulatory	For research purposes only
-------------------	----------------------------

North & South Tel: +1 800 265 7376

America Fax: +1 919 878 3751

Email: antibody_sales_us@bio-rad.com

Worldwide

Tel: +44 (0)1865 852 700

Fax: +44 (0)1865 852 739

Email: antibody_sales_uk@bio-rad.com

Europe

Tel: +49 (0) 89 8090 95 21

Fax: +49 (0) 89 8090 95 50

Email: antibody_sales_de@bio-rad.com

'M301164:170106'

Printed on 01 Aug 2018

© 2018 Bio-Rad Laboratories Inc | [Legal](#) | [Imprint](#)